

# Z-300

## Storage chamber





## **Technical concept**

Refrigerated preservation system for semen doses that prevents thermal shock and keeps doses under optimal conditions till the moment of insemination. It is specially designed and built to keep semen doses at a constant temperature.

#### j}

#### **TECHNICAL SPECIFICATIONS**

Supply voltage: 220-240V Frequency: 50Hz

Refrigerant fluid: R-600a / 45g

Nominal power: 225W

Max current at compressor start-up: 5A

Nominal useable volume: 256 l

Optimal no. of doses: 520 doses

Temp. measurement resolution: 0.1°C

Temperature range: 15°C to 40°C +/-0.5°C

External dimensions: 550x612x1250 (height) mm

Unladen weight: 47 Kg

## **Advantages**

- The Z-300 is equipped with a based digital regulator, which controls the temperature inside the container.
- It also has an alarm system which allows the operator to check whether a serious deviation in temperature has occurred at any time that would affect the preservation of the serious.
- To ensure the temperature is an even as possible inside the container, the machine has a forced-ventilation system that guarantees constant recirculation of air.
- The container's interior is designed to enable the automatic collection and evaporation of condensates.
- Designed and adapted to satisfy the highest demands in terms of reliability and safety during its whole useful life.
- Energy classification A+.
- 5 shelves.
- Quiet operation.









**Our recommendations** for a proper preservation and storage of seminal doses:

- Check that this preservation device maintains the homogeneous inner temperature. For this, we recommend a constant indoor air flow recirculation.
- >>> Place the seminal doses evenly inside the device.
- >>> Place the machine in non-closed places to ease breathing of the engine.
- >>> Remember that continually opening the equipment affects its performance and consumption.
- >> Keep doses between 15.5 ° C and 16.5 ° C
- Avoid temperature fluctuations during the storage period. The critical temperature at which irreversible damage to the sperm membrane occurs is below 12 ° C
- Do not expose the doses for long periods of time to direct light.
- >> Do not rotate doses during storage.
- >> If in doubt, contact our technical service.





